

Declaration of Conformity

Manufacturer: Harvard Apparatus
Address: 84 October Hill Road, Holliston, Massachusetts 01746, USA

We herewith declare that the following products:

Product Name: Infusion Pump
Models: 70-2219 70-2225
70-2220 70-2226

Are in conformity of the following applicable European regulations and directives:

2014/35/EU Low Voltage directive (LVD)
2014/30/EU Electromagnetic Compatibility directive (EMC)
2012/19/EU Waste electrical and electronic equipment directive (WEEE)²
2011/65/EU Restriction on the use of certain hazardous substances directive (RoHS)³

Standards used to demonstrate conformity include:

EN 61010-1:2010 Safety requirements for electrical equipment for measurement, control and laboratory use, General Requirements
EN 61326-1:2013 Electrical equipment for measurement, control and laboratory use –EMC requirements²
EN 50581:2012 Assessment of products with respect to RoHS³

EMC and Safety testing performed by Intertek Testing Services.
Listing Number 9801781
Test reports: 101450192BOX-003, 3025152 US/CAN, 3125828LTT-003
Conforms to UL61010-1:2012
Certified to CSA C22.2#61010-1:2012



¹Harvard Bioscience hold an ISO 9001:2015 certificate BSI FM 629138 for its Quality Management System
²Classified and tested as class A equipment in accordance with CISPR 11 definition in a basic electromagnetic environment.
³Classified as belonging to equipment categories 8 or 9

I, the undersigned, hereby declare that the equipment specified is in conformity with the relevant harmonised Union legislation. Signed for and on behalf of BTX at Holliston, USA



Signed:

Mark Davis
Harvard Bioscience, Director of
Engineering

Date:
November 9, 2018